

Date of Document: 14/11/2025
Clinic Name: Inam Muhammad
Clinic Address: 2a Cyclamen Avenue Altona North
Clinic Phone: 0449 581 689
Clinic Contact: Inam Muhammad

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clippers- Own	20	14/11/2025	2251114-1	13/11/2026
Scalpel	8	14/11/2025	2251114-1	13/11/2026

Total items: 28 **Batch** **Sets**
Clean boxes Supplied: 1 2251114-1 28
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:



Notes:

Getinge Assured
 Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 240101

