

Date of Document: 19/1/2026
Clinic Name: Molemap Brighton
Clinic Address: 14/3 Male St Brighton 3186
Clinic Phone: 1800 665 362
Clinic Contact: Sophie

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Suture Set	6	19/1/2026	260119-4	18/1/2027
Forcep- Ribbed	5	19/1/2026	260119-4	18/1/2027
Forcep- Toothed	1	19/1/2026	260119-4	18/1/2027
Small Scissors	1	19/1/2026	260119-4	18/1/2027

Total items: 13 **Batch** **Sets**
Clean boxes Supplied: 1 260119-4 13
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:





Getinge
Assured
 Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Notes:

Sterilised By PodMed Supplies
 STERILISATION DATE: 19/01/2026
 BATCH NUMBER: 260119-4
 EXPIRATION DATE: 18/01/2027



STERILE 
 DO NOT USE IF PACK IS DAMAGED