PODMED SUPPLIES

Date of Document: 13/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Ozcare Hervey Bay Aged Care Facility - 35 Davis Dr, Kawungan QLD 4655
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	20	13/1/2026	2260113-1	12/1/2027

Total items:	20	**Batch**	**Sets**
Clean boxes Supplied:	1	2260113-1	20
Dirty Boxes supplied:	1		

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 13/01/26 AUTOCLAVE 2
BROWN STERILIZED
2260113-1 BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Sterilised By PodMed Supplies
STERILISATION DATE: 13/01/2026
BATCH NUMBER: 2260113-1
EXPIRATION DATE: 12/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED

Notes: