PODMED SUPPLIES

Date of Document: **14/1/2026**
Clinic Name: Australian Health Professionals
Clinic Address: C/O 604/31 Crown St, Wollongong NSW 2500
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	21	14/1/2026	260114-2	13/1/2027

Total items: 21
Clean boxes Supplied: 1
Dirty Boxes supplied: 1

Batch 260114-2
Sets 21

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 14/01/26 AUTOCLAVE 1
BROWN - STERILIZED
260114-2 BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 14/01/2026
BATCH NUMBER: 260114-2
EXPIRATION DATE: 13/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED