PODMED SUPPLIES

Date of Document: **26/11/2025**
Clinic Name: Stride Footcare
Clinic Address: 3/24 Beesley St, West End QLD 4101
Clinic Phone: 0469 197 978
Clinic Contact: Cristina

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Packs- Diamond/ cerami	20	26/11/2025	251126-2	25/11/2026

Total items: 20 **Batch** **Sets**
Clean boxes Supplied: 1 0 20
Dirty Boxes supplied: 1

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 26/11/25 AUTOCLAVE 1
BROWN STERILIZED
251126-2
BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 240101

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 26/11/2025
BATCH NUMBER: 251126-2
EXPIRATION DATE: 25/11/2026
STERILE
DO NOT USE IF PACK IS DAMAGED