

Date of Document: 8/12/2025
Clinic Name: Molemap Brighton
Clinic Address: 14/3 Male St Brighton 3186
Clinic Phone: 1800 665 362
Clinic Contact: Sophie


Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Suture Set	4	8/12/2025	251208-1	7/12/2026
Forcep- Ribbed	3	8/12/2025	251208-1	7/12/2026

Total items: 7 **Batch** 251208-1 **Sets** 7
Clean boxes Supplied: 1
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H


Signed:


Getinge Assured
 Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 240101

Notes:

Sterilised By PodMed Supplies
 STERILISATION DATE: 08/12/2025
 BATCH NUMBER: 251208-1
 EXPIRATION DATE: 08/12/2026



STERILE 

DO NOT USE IF PACK IS DAMAGED