

Date of Document: 7/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Tumut - 112 Lambie St, Tumut NSW 2720
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	21	7/1/2026	260107-1	6/1/2027

Total items: 21 **Batch** **Sets** 21
Clean boxes Supplied: 1 260107-1
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:



Notes:

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Sterilised By PodMed Supplies
 STERILISATION DATE: 07/01/2026
 BATCH NUMBER: 260107-1
 EXPIRATION DATE: 06/01/2027
STERILE 
 DO NOT USE IF PACK IS DAMAGED

