

Date of Document: 21/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Pacifica Nambucca Heads
 11 Short Street Nambucca Heads
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clippers + file	20	21/1/2026	2260121-1	20/1/2027

Total items: 20 **Batch** **Sets**
Clean boxes Supplied: 1 2260121-1 20
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:



Getinge Assured
 Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Notes:

Sterilised By PodMed Supplies

STERILISATION DATE: 21/01/2026

BATCH NUMBER: 2260121-1

EXPIRATION DATE: 20/01/2027

STERILE

DO NOT USE IF PACK IS DAMAGED

