

Date of Document: 30/12/2025
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Wontama 30-38 Byng St Orange NSW
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	21	30/12/2025	2251230-2	29/12/2026

Total items: 21 **Batch** **Sets**
Clean boxes Supplied: 1 2251230-2 21
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:



Notes:

Getinge
Assured
 Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Sterilised By PodMed Supplies
 STERILISATION DATE: 30/12/2025
 BATCH NUMBER: 2251230-2
 EXPIRATION DATE: 29/12/2026
STERILE
 DO NOT USE IF PACK IS DAMAGED

