PODMED SUPPLIES

Date of Document: **30/1/2026**
Clinic Name: Stride Podiatry
Clinic Address: 3/24 Beesley Street West End QLD 4101
Clinic Phone: 0.46919798
Clinic Contact: Cristina

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Pack- Med Diamond Bur	10	30/1/2026	260130-2	29/1/2027
Podiatry Treatment Pack- Med Ceramic Burr	10	30/1/2026	260130-2	29/1/2027

Total items:	20	**Batch**	**Sets**
Clean boxes Supplied:	1	260130-2	20
Dirty Boxes supplied:	1		

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE | AUTOCLAVE
30/01/26 1
BROWNE STERILINER
260130-2
BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 30/01/2026
BATCH NUMBER: 260130-2
EXPIRATION DATE: 29/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED