

**Date of Document:** 22/9/2025  
**Clinic Name:** The Podiatry Co. Geelong  
**Clinic Address:** The Welfare Hub - 245 Roslyn Rd, Highton, 3216  
**Clinic Phone:** 0411 292 231  
**Clinic Contact:** Alice

**Sterile instrument order:**

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Sets	20	22/9/2025	250922-2	21/9/2026

**Total items:** 20      **Batch** 250922-2      **Sets** 20  
**Clean boxes Supplied:** 1  
**Dirty Boxes supplied:** 1

Sterile compliant on autoclave      YES / NO  
 Sterile compliant on pack      YES / NO  
 Sterile compliant on class 6 emulating indicator      YES / NO  
 Sterile compliant on indicator label      YES / NO

**Certified By:** Chris H

**Signed:**



**Notes:**

CYCLE DATE	AUTOCLAVE
22/09/25	1
250922-2	
BATCH NUMBER	

  

**Getinge**  
**Assured**  
 Emulating Indicator 5.3

**STEAM** ISO 11140-1 Type 6  
 5.3 Minutes @ 134°C  
 15 Minutes @ 121°C  
**LOT** 240101