

**Date of Document:** 16/1/2026  
**Clinic Name:** A2Z Home Podiatry  
**Clinic Address:** 12 Arabella Circuit, Point Cook VIC 3030  
**Clinic Phone:** 0427035998  
**Clinic Contact:** Albin

**Sterile instrument order:**

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Home Home Visist Set	20	16/1/2026	2260116-3	15/1/2027

**Total items:** 20      **Batch**      **Sets**  
**Clean boxes Supplied:** 1      2260116-3      20  
**Dirty Boxes supplied:** 1

Sterile compliant on autoclave  
 Sterile compliant on pack  
 Sterile compliant on class 6 emulating indicator  
 Sterile compliant on indicator label

YES / NO  
 YES / NO  
 YES / NO  
 YES / NO



**Certified By:** Chris H

**Signed:**



**Getinge Assured**  
 Emulating Indicator 5.3

**STEAM** ISO 11140-1 Type 6  
 5.3 Minutes @ 134°C  
 15 Minutes @ 121°C  
**LOT** 241001

**Notes:**

Sterilised By PodMed Supplies  
 STERILISATION DATE: 16/01/2026  
 BATCH NUMBER: 2260116-3  
 EXPIRATION DATE: 15/01/2027

  
**STERILE**   
 DO NOT USE IF PACK IS DAMAGED