PODMED SUPPLIES

Date of Document: 21/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Pacifica Nambucca Heads
11 Short Street Nambucca Heads
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clippers + file	14	21/1/2026	260121-3	20/1/2027

Total items: 14
Clean boxes Supplied: 1
Dirty Boxes supplied: 1

Batch	Sets
260121-3	14

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 21/01/26 AUTOCLAVE 1
BROWN STERILIZER
260121-3 BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 21/01/2026
BATCH NUMBER: 260121-3
EXPIRATION DATE: 20/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED