PODMED SUPPLIES

Date of Document: 16/1/2026
Clinic Name: Jia Ing
Clinic Address: 104/ 26 Hood St Subiaco WA 6008
Clinic Phone: 0420475347
Clinic Contact: Jia Ing

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podaitry Treatment Pack	20	16/1/2026	260116-4	15/1/2027

Total items: 20
Clean boxes Supplied: 1
Dirty Boxes supplied: 1

Batch 260116-4
Sets 20

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE | AUTOCLAVE
16/01/26 1
BROWN STERILIZED
260116-4
BATCH NUMBER

Certified By: Chris H

Getinge Assured
Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 16/01/2026
BATCH NUMBER: 260116-4
EXPIRATION DATE: 16/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED