

**Date of Document:** 5/12/2025  
**Clinic Name:** Molemap Carlton  
**Clinic Address:** 237a Faraday St, Carlton VIC 3053  
**Clinic Phone:** 1800 665 362  
**Clinic Contact:** Matthew


**Sterile instrument order:**

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Suture Sets	9	5/12/2025	251205-2	4/12/2026
Needle Holder Large	3	5/12/2025	251205-2	4/12/2026
Toothed Forceps	6	5/12/2025	251205-2	4/12/2026
Ribbed Forceps	4	5/12/2025	251205-2	4/12/2026
Needle Holder Small	1	5/12/2025	251205-2	4/12/2026
Curved Scissors	1	5/12/2025	251205-2	4/12/2026

**Total items:** 24      **Batch**      **Sets**  
**Clean boxes Supplied:** 1      251205-2      24  
**Dirty Boxes supplied:** 1

Sterile compliant on autoclave  
 Sterile compliant on pack  
 Sterile compliant on class 6 emulating indicator  
 Sterile compliant on indicator label

YES / NO  
 YES / NO  
 YES / NO  
 YES / NO




**Certified By:** Chris H

**Signed:**



**Getinge Assured**  
 Emulating Indicator 5.3


**STEAM** ISO 11140-1 Type 6  
 5.3 Minutes @ 134°C  
 15 Minutes @ 121°C  
**LOT** 240101



**Notes:**

Sterilised By PodMed Supplies

STERILISATION DATE: 05/12/2025  
 BATCH NUMBER: 251205-2  
 EXPIRATION DATE: 4/12/2026



  
 DO NOT USE IF PACK IS DAMAGED