

Date of Document: 19/3/2025
Clinic Name: Stride Podiatry
Clinic Address: 66 McLennan St Albion 4010 QLD
Clinic Phone: 0.46919798
Clinic Contact: Cristina

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Pack- Med TC Burr	20	19/3/2025	250319:3	18/3/2026

Total items: 20
Clean boxes Supplied: 1
Dirty Boxes supplied: 1

Batch 250319:3 **Sets** 20

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO

Certified By: Chris H

Signed:



Notes:

CYCLE DATE | AUTOCLAVE
 19/03/25 | 1
 BROWN-STERILIZED
 250319-3
 BATCH NUMBER

Emulating Indicator
 134°C - 3,5 min. / 121°C - 15 min.

[REF] 3FSKS645818
 EN-ISO 11140-1
 Type 6 indicator
 [LOT] 1318267
 2023/05 2026/05
 Manufactured in The Netherlands.
 www.interster.nl
INTERSTER
 A-19



STEAM

Minimum reference colour

Sterilizing conditions turn indicator from light blue to dark.



Assembled by: CH
 Date: 19/3