

Date of Document: 14/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: C/O 604/31 Crown St, Wollongong NSW 2500
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	21	14/1/2026	2260114-2	13/1/2027

Total items: 21 **Batch** **Sets**
Clean boxes Supplied: 1 2260114-2 21
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO

CYCLE DATE AUTOCLAVE
 14/01/26 2
 BROWN-STERILIZED
 2260114-2
 BATCH NUMBER

Certified By: Chris H
Signed:

Notes:

Getinge Assured
 Emulating Indicator 5.3 **STEAM** ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C **LOT** 241001

Sterilised By PodMed Supplies
 STERILISATION DATE: 14/01/2026
 BATCH NUMBER: 2260114-2
 EXPIRATION DATE: 13/01/2027
STERILE 
 DO NOT USE IF PACK IS DAMAGED

