

Date of Document: 13/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Ozcare Hervey Bay Aged Care Facility -
 35 Davis Dr, Kawungan QLD 4655
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	13	13/1/2026	260113-3	12/1/2027
Scalpel Handle	6	13/1/2026	260113-3	12/1/2027
Blacks File	3	13/1/2026	260113-3	12/1/2027
Scalpel Handle	4	15/12/2025	2251215-3	14/12/2026
Blacks File	2	15/12/2025	2251215-3	14/12/2026

Total items:	28	Batch	Sets
Clean boxes Supplied:	1	260113-3	21
Dirty Boxes supplied:	1	2251215-3	6

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO


Certified By: Chris H
Signed:

Notes:

Getinge Assured
 Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Sterilised By PodMed Supplies
 STERILISATION DATE: 13/01/2026
 BATCH NUMBER: 260113-3
 EXPIRATION DATE: 12/01/2027

STERILE

DO NOT USE IF PACK IS DAMAGED

