PODMED SUPPLIES

Date of Document: **30/12/2025**
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Wontama 30-38 Byng St Orange NSW
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clipper + File	21	30/12/2025	2251230-1	29/12/2026

Total items:	21	**Batch**	**Sets**
Clean boxes Supplied:	1	2251230-1	21
Dirty Boxes supplied:	1		

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 30/12/25 AUTOCLAVE 2
BROWN STERILIZER
2251230-1 BATCH NUMBER

Certified By: Chris H

Signed:

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 30/12/2025
BATCH NUMBER: 2251230-1
EXPIRATION DATE: 29/12/2026
STERILE
DO NOT USE IF PACK IS DAMAGED