PODMED SUPPLIES

STERILISATION COMPLIANCE DOCUMENTATION

Sterilised in accordance with AS 5369:2023

Date of Document: 16/02/2026

Clinic Name: Inam Muhammad

Clinic Address: 2a Cyclamen Ave Altona North VIC 3025

Clinic Phone: 0429 581 689

Clinic Contact: Inam Muhammad

Sterilising Batch Number: 260216-2

Sterilising Batch Date: 16/02/2026

Sterilising Batch Expiry: 15/02/2027

Sterilising Batch Contents

Instruments	Count
Podiatry clippers	20
Scalpel handle	3
Blacks file	1
Total Count:	24

Clean Boxes Supplied: 1

Boxes for used instruments: 1

Compliance Checked *(ticked if successful)*:

- ✓ Sterile compliant on autoclave
- ✓ Sterile compliant on pack
- ✓ Sterile compliant on class 6 emulating indictor
- ✓ Sterile compliant on Printex indicator label

Certified By: Chris H

Signed:



Compliance Labels & Indicators:

Batch: 260216-2

Notes