

Date of Document: 30/12/2025
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Wontama 30-38 Byng St Orange NSW
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Scalpel Handle	40	30/12/2025	251230-4	29/12/2026
Blacks File	30	30/12/2025	251230-3	29/12/2026

Total items: 70 **Batch** **Sets**
Clean boxes Supplied: 1 251230-4 40
Dirty Boxes supplied: 1 251230-3 30

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO



Certified By: Chris H

Signed:



Notes:

Getinge Assured
Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Getinge Assured
Emulating Indicator 5.3

STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Sterilised By PodMed Supplies

STERILISATION DATE: 30/12/2025
 BATCH NUMBER: 251230-4
 EXPIRATION DATE: 29/12/2026

STERILE 

DO NOT USE IF PACK IS DAMAGED

Sterilised By PodMed Supplies

STERILISATION DATE: 30/12/2025
 BATCH NUMBER: 251230-3
 EXPIRATION DATE: 29/12/2026

STERILE 

DO NOT USE IF PACK IS DAMAGED