

Date of Document: 20/1/2026
Clinic Name: Camberwell SSM
Clinic Address: 1431 Toorak Rd Camberwell 3124
Clinic Phone: 03 98891078
Clinic Contact: Vaughan Ackland

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Pack- PodMed	20	20/1/2026	2260120-3	19/1/2027

Total items: 20 **Batch** **Sets**
Clean boxes Supplied: 1 2260120-3 20
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed:



Getinge Assured
 Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
LOT 241001

Notes:

Sterilised By PodMed Supplies
 STERILISATION DATE: 20/01/2026
 BATCH NUMBER: 2260120-3
 EXPIRATION DATE: 19/01/2027
STERILE
 DO NOT USE IF PACK IS DAMAGED

