

Date of Document: 8/10/2025
Clinic Name: Stride Footcare
Clinic Address: 3/24 Beesley St, West End QLD 4101
Clinic Phone: 0469 197 978
Clinic Contact: Cristina

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
General Treatment Pack- Diamond Burr	10	8/10/2025	251008-2	7/10/2026
General Treatment Pack- Ceramic Burr	10	8/10/2025	251008-2	7/10/2026
Scalpel Handles	6	8/10/2025	251008-2	7/10/2026

Total items: 26 **Batch** **Sets**
Clean boxes Supplied: 1 251008-2 20
Dirty Boxes supplied: 1

Sterile compliant on autoclave YES / NO
 Sterile compliant on pack YES / NO
 Sterile compliant on class 6 emulating indicator YES / NO
 Sterile compliant on indicator label YES / NO



Certified By: Chris H

Signed:



Getinge Assured
 Emulating Indicator 5.3 **STEAM** ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C **LOT** 240101

Notes: