

Date of Document: 23/1/2026
Clinic Name: Stride Podiatry
Clinic Address: 3/24 Beesley Street West End QLD 4101
Clinic Phone: 0.46919798
Clinic Contact: Cristina


Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Pack- Med Diamond Burr	10	23/1/2026	2260123-2	22/1/2027
Podiatry Treatment Pack- Med Ceramic Burr	10	23/1/2026	2260123-2	22/1/2027

Total items: 20 **Batch** **Sets** 20
Clean boxes Supplied: 1 2260123-2
Dirty Boxes supplied: 1

Sterile compliant on autoclave
 Sterile compliant on pack
 Sterile compliant on class 6 emulating indicator
 Sterile compliant on indicator label

YES / NO
 YES / NO
 YES / NO
 YES / NO



Certified By: Chris H

Signed: 

Getinge Assured **STEAM ISO 11140-1 Type 6**
Emulating Indicator 5.3 5.3 Minutes @ 134°C
 15 Minutes @ 121°C
 LOT 241001

Notes:

Sterilised By PodMed Supplies
 STERILISATION DATE: 23/01/2026
 BATCH NUMBER: 2260123-2
 EXPIRATION DATE: 22/01/2027

STERILE 
 DO NOT USE IF PACK IS DAMAGED