PODMED SUPPLIES

Date of Document: 21/1/2026
Clinic Name: Australian Health Professionals
Clinic Address: Uniting Pacifica Nambucca Heads
11 Short Street Nambucca Heads
Clinic Phone: 0420 241 900
Clinic Contact: Vivian

Sterile insturment order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Clippers + file	20	21/1/2026	260121-1	20/1/2027

Total items:	20	**Batch**	**Sets**
Clean boxes Supplied:	1	260121-1	20
Dirty Boxes supplied:	1		

Sterile compliant on autoclave YES / NO
Sterile compliant on pack YES / NO
Sterile compliant on class 6 emulating indicator YES / NO
Sterile compliant on indicator label YES / NO

CYCLE DATE 21/01/26 AUTOCLAVE 1
STERILIZED
260121-1 BATCH NUMBER

Certified By: Chris H

Getinge Assured Emulating Indicator 5.3
STEAM ISO 11140-1 Type 6
5.3 Minutes @ 134°C
15 Minutes @ 121°C
LOT 241001

Signed:

Notes:

Sterilised By PodMed Supplies
STERILISATION DATE: 21/01/2026
BATCH NUMBER: 260121-1
EXPIRATION DATE: 20/01/2027
STERILE
DO NOT USE IF PACK IS DAMAGED