

**Date of Document:** 7/1/2026  
**Clinic Name:** Australian Health Professionals  
**Clinic Address:** Uniting Tumut - 112 Lambie St, Tumut NSW 2720  
**Clinic Phone:** 0420 241 900  
**Clinic Contact:** Vivian

**Sterile instrument order:**

| Item           | Quantity | Sterilisation Date | Tracking number | Expiry Date |
|----------------|----------|--------------------|-----------------|-------------|
| Clipper + File | 22       | 7/1/2026           | 260107-4        | 6/1/2027    |
|                |          |                    |                 |             |
|                |          |                    |                 |             |
|                |          |                    |                 |             |
|                |          |                    |                 |             |
|                |          |                    |                 |             |
|                |          |                    |                 |             |

**Total items:** 22      **Batch**      **Sets**  
**Clean boxes Supplied:** 1      260107-4      22  
**Dirty Boxes supplied:** 1

Sterile compliant on autoclave  
 Sterile compliant on pack  
 Sterile compliant on class 6 emulating indicator  
 Sterile compliant on indicator label

YES / NO  
 YES / NO  
 YES / NO  
 YES / NO

CYCLE DATE      AUTOCLAVE  
 07/01/26      1  
 260107-4  
 BATCH NUMBER

**Certified By:** Chris H

**Signed:**



**Getinge Assured**  
**Emulating Indicator 5.3**  
**STEAM ISO 11140-1 Type 6**  
**5.3 Minutes @ 134°C**  
**15 Minutes @ 121°C**  
**LOT 241001**

**Notes:**

Sterilised By PodMed Supplies  
 STERILISATION DATE: 07/01/2026  
 BATCH NUMBER: 260107-4  
 EXPIRATION DATE: 06/01/2027  
 STERILE   
 DO NOT USE IF PACK IS DAMAGED

