

**Date of Document:** 12/12/2025  
**Clinic Name:** Molemap Carlton  
**Clinic Address:** 237a Faraday St, Carlton VIC 3053  
**Clinic Phone:** 1800 665 362  
**Clinic Contact:** Matthew

**Sterile instrument order:**

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Suture Sets	4	12/12/2025	2251212-3	11/12/2026
Ribbed Forceps	6	12/12/2025	2251212-3	11/12/2026
Needle Holder Small	1	12/12/2025	2251212-3	11/12/2026

**Total items:** 11      **Batch** 2251212-3      **Sets** 11  
**Clean boxes Supplied:** 1  
**Dirty Boxes supplied:** 1

Sterile compliant on autoclave  
 Sterile compliant on pack  
 Sterile compliant on class 6 emulating indicator  
 Sterile compliant on indicator label

YES / NO  
 YES / NO  
 YES / NO  
 YES / NO

CYCLE DATE 12/12/25      AUTOCLAVE 2  
 2251212-3  
 BATCH NUMBER

**Certified By:** Chris H

**Signed:**



**Notes:**

**Getinge Assured**  
 Emulating Indicator 5.3      **STEAM** ISO 11140-1 Type 6  
 5.3 Minutes @ 134°C  
 15 Minutes @ 121°C      **LOT** 240101

Sterilised By PodMed Supplies  
 STERILISATION DATE: 12/12/2025  
 BATCH NUMBER: 2251212-3  
 EXPIRATION DATE: 11/12/2026  
**STERILE**   
 DO NOT USE IF PACK IS DAMAGED

