

Date of Document: 21/11/2025
Clinic Name: Stride Footcare
Clinic Address: 3/24 Beesley St, West End QLD 4101
Clinic Phone: 0469 197 978
Clinic Contact: Cristina

Sterile instrument order:

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Packs- Diamond	10	21/11/2025	2251121-2	20/11/2026
Podiatry Treatment Packs- Ceramic	10	21/11/2025	2251121-2	20/11/2026
Galipot	1	21/11/2025	2251121-2	20/11/2026
Nail Surgery Pack	1	21/11/2025	2251121-2	20/11/2026
Curette	1	21/11/2025	2251121-2	20/11/2026

Total items: 23 **Batch** **Sets**
Clean boxes Supplied: 1 2251121-2 23
Dirty Boxes supplied: 1

Sterile compliant on autoclave YES / NO
 Sterile compliant on pack YES / NO
 Sterile compliant on class 6 emulating indicator YES / NO
 Sterile compliant on indicator label YES / NO

Certified By: Chris H

Signed:




Getinge Assured
 Emulating Indicator 5.3 **STEAM** ISO 11140-1 Type 6
 5.3 Minutes @ 134°C
 15 Minutes @ 121°C **LOT** 240101

Notes:

Sterilised By PodMed Supplies
 STERILISATION DATE: 21/11/2025
 BATCH NUMBER: 2251121-2
 EXPIRATION DATE: 20/11/2026
STERILE 
 DO NOT USE IF PACK IS DAMAGED

