

**Date of Document:** 23/1/2026  
**Clinic Name:** The Podiatry Co. Geelong  
**Clinic Address:** The Welfare Hub - 245 Roslyn Rd, Highton, 3216  
**Clinic Phone:** 0411 292 231  
**Clinic Contact:** Alice

**Sterile instrument order:**

Item	Quantity	Sterilisation Date	Tracking number	Expiry Date
Podiatry Treatment Sets	20	23/1/2026	260123-1	22/1/2027

<b>Total items:</b>	20	<b>Batch</b>	260123-1	<b>Sets</b>	20
<b>Clean boxes Supplied:</b>	1				
<b>Dirty Boxes supplied:</b>	1				

Sterile compliant on autoclave  
 Sterile compliant on pack  
 Sterile compliant on class 6 emulating indicator  
 Sterile compliant on indicator label

YES / NO  
 YES / NO  
 YES / NO  
 YES / NO

CYCLE DATE 23/01/26 AUTOCLAVE  
 260123-1  
 BATCH NUMBER

**Certified By:** Chris H

**Signed:**



**Notes:**

**Getinge Assured**  
 Emulating Indicator 5.3  
**STEAM** ISO 11140-1 Type 6  
 5.3 Minutes @ 134°C  
 15 Minutes @ 121°C  
**LOT** 241001

Sterilised By PodMed Supplies  
 STERILISATION DATE: 23/01/2026  
 BATCH NUMBER: 260123-1  
 EXPIRATION DATE: 22/01/2027  
**STERILE**  
 DO NOT USE IF PACK IS DAMAGED

